

Laboratory	ASKIB Engineers Private Limited, 86 D, Dr. Suresh Sarkar Road, Kolkata, West Bengal		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	20.11.2014
Certificate Number	C-1157	Valid Until	19.11.2016
Last Amended on	-	Page	1 of 5

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>SOURCE</u>			
1. DC VOLTAGE^{\$}	1 mV	1.443 %	Using ZEAL Multi Function Calibrator By Direct Method
	100 mV	0.24 %	
	10 V	0.24 %	
	1000 V	0.24 %	
2. DC CURRENT^{\$}	1 mA	0.64 %	Using ZEAL Multi Function Calibrator By Direct Method
	100 mA	0.25 %	
	10 A	0.34 %	
3. AC VOLTAGE^{\$}	50 Hz		Using ZEAL Multi Function Calibrator By Direct Method
	10 mV	2.31 %	
	100 mV	0.47 %	
	10 V	0.435 %	
4. AC CURRENT^{\$}	50 Hz		Using ZEAL Multi Function Calibrator By Direct Method
	100 mA	0.40 %	
	10 A	1.2 %	
5. RESISTANCE^{\$}	1 Ω	4.06 %	Using Decade Resistance Box By Direct Method
	10 Ω	1.183 %	
	1 G Ω	3.46%	
6. FREQUENCY^{\$}	45 Hz	0.60 %	Using ZEAL Multi Function Calibrator By Direct Method
	990 Hz	0.213 %	

Avijit Das
Program Manager

Sangeeta Kunwar
Convenor

Laboratory	ASKIB Engineers Private Limited, 86 D, Dr. Suresh Sarkar Road, Kolkata, West Bengal		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	20.11.2014
Certificate Number	C-1157	Valid Until	19.11.2016
Last Amended on	-	Page	2 of 5

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
7. CAPACITANCE^s	1 kHz		
	1 nF	5.39 %	Using Decade Capacitance Box By Direct Method
	10 μ F	3.70%	
8. INDUCTANCE^s	1 kHz		
	1 mH	4.00 %	Using Decade Capacitance Box By Direct Method
	10 H	4.00 %	
9. Temperature Simulation[#] (Controller/ Indicator/ Recorder)			
RTD	-200°C	0.16 %	Using Process Calibrator 754 By Direct Method
	800°C	0.36%	
T – Type T/C	-200°C	0.87 %	
	400°C	0.41 %	
S – Type T/C	0°C	1.58°C	
	1700°C	1.58°C	
J – Type T/C	-200°C	0.49 %	
	1200°C	0.6 %	
N – Type T/C	-200°C	0.85 %	
	1300°C	0.5 %	
K – Type T/C	-200°C	0.62 %	
	1300°C	0.5 %	
R – Type T/C	0°C	1.5°C	
	1700°C	1.5°C	
E – Type T/C	-200°C	0.7 %	
	1000°C	0.43 %	
B – Type T/C	600°C	1.3 %	
	1800°C	1.0 %	

Avijit Das
Program Manager

Sangeeta Kunwar
Convenor

Laboratory	ASKIB Engineers Private Limited, 86 D, Dr. Suresh Sarkar Road, Kolkata, West Bengal		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	20.11.2014
Certificate Number	C-1157	Valid Until	19.11.2016
Last Amended on	-	Page	3 of 5

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>MEASURE</u>			
1. DC VOLTAGE^{\$}	100 μ V	4.05 %	Using Digital Multimeter 6.5 By Direct Method
	100 mV	0.20 %	
	10 V	0.08 %	
	1000 V	0.08 %	
2. DC CURRENT^{\$}	100 μ A	0.12 %	Using Digital Multimeter 6.5 By Direct Method
	100 mA	0.50 %	
	10 A	0.33 %	
3. AC VOLTAGE^{\$}	50 Hz		Using Digital Multimeter 6.5 By Direct Method
	100 mV	0.23 %	
	10 V	0.4 %	
	1000 V	0.2 %	
4. AC CURRENT^{\$}	50 Hz		Using Digital Multimeter 6.5 By Direct Method
	20 mA	0.7 %	
	100 mA	0.62 %	
5. RESISTANCE^{\$}	1.25 Ω	0.76 %	Using Digital Multimeter 6.5 By Direct Method
	10 Ω	0.40 %	
	1 M Ω	0.17 %	
	100 M Ω	1.11 %	
6. FREQUENCY^{\$}	10 Hz	0.11 %	Using Digital Multimeter 6.5 By Direct Method
	1 MHz	0.082 %	

Avijit Das
Program Manager

Sangeeta Kunwar
Convenor

Laboratory	ASKIB Engineers Private Limited, 86 D, Dr. Suresh Sarkar Road, Kolkata, West Bengal		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	20.11.2014
Certificate Number	C-1157	Valid Until	19.11.2016
Last Amended on	-	Page	4 of 5

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
7. CAPACITANCE	10 nF 10 μ F	3.0 % 3.01 %	Using Digital Multimeter 6.5 By Direct Method
8. STOP WATCH ^{\$} (Digital/Analog)	5 min to 24 hr	3.55 %	Using Digital Stop Watch By Comparison Method
9. AC HIGH VOLTAGE [#]	1 kV to 25 kV	6.11 %	Using H.V probe with DMM
10 Temperature Simulation [#] (Temperature Source/ Temperature Calibrator)			
RTD	-200°C 800°C	0.19 0.32	Process Calibrator 754 By Direct Method
T – Type T/C	-200°C 400°C	1.00 0.41	
S – Type T/C	0°C 1700°C	1.67°C 1.67°C	
J – Type T/C	-200°C 1200°C	0.78 0.73	
N – Type T/C	-200°C 1300°C	1.25 0.78	
K – Type T/C	-200°C 1300°C	0.91 0.88	
R – Type T/C	100°C 1700°C	1.56°C 1.56°C	

Avijit Das
Program Manager

Sangeeta Kunwar
Convenor

Laboratory	ASKIB Engineers Private Limited, 86 D, Dr. Suresh Sarkar Road, Kolkata, West Bengal		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electro-Technical Calibration	Issue Date	20.11.2014
Certificate Number	C-1157	Valid Until	19.11.2016
Last Amended on	-	Page	5 of 5

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
E – Type T/C	-200°C	1.62	
	1000°C	0.59	
B– Type T/C	600 °C	1.8	
	1800 °C	1.4	

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%.

\$ Only in Permanent Laboratory

* Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Avijit Das
Program Manager

Sangeeta Kunwar
Convenor